

## Scope of Accreditation For RT Precision Measurement Inc.

2679 Howard Avenue  
Windsor, Ontario N8X 3X2  
Craig Renaud & Byron Tanner  
519-969-7785

In recognition of a successful assessment to ISO/IEC 17025:1999, accreditation is granted to **RT Precision Measurement Inc.**, to perform **Dimensional Inspection** in the following parameters:


Accreditation granted through: **June 30, 2007**

### Dimensional Inspection

Inspection Parameter/Equipment	Range	Best Measurement Capability (+/-)	Remarks
Dimensional Inspection  Volumetric Accuracy	X = 0 mm to 2500 mm Y = 0 mm to 1500 mm Z = 0 mm to 1000 mm	(15 + 4L) μm	DEA Omega Coordinate Measuring Machine and Utilized as Reference Standards

Notes:

- 1) Laboratory offers calibration services at the laboratory's own facilities and at the client or other agreed upon facilities.
- 2) Best uncertainties represent expanded uncertainties at approximately the 95% confidence level using a coverage factor of k=2.

Approved by:   
R. Douglas Leonard Jr.  
Chief Technical Officer

Date: June 30, 2004

Issued: 6/30/04